



#### DESCRIPTION:

Transducer interface cable with blue collared male connector. Cable connects a disposable pressure transducer with the MPS<sup>®</sup> console.

#### INDICATION FOR USE

The transducer interface cable is indicated to be used as the electronic interface between the MPS console and a disposable pressure transducer for purpose of monitoring a defined pressure.

#### CONTRAINDICATION

This product is designed and warranted for use with the Quest MPS cardioplegia delivery system only. Use with other systems is contraindicated. This product is not designed, sold, or intended for use except as indicated.

#### WARNINGS

Read instructions carefully. The attending clinician is solely responsible for the setup and use of this device.

Use cable to interface disposable pressure transducer with the MPS console.

The transducer will not accurately measure pressures above 300 mmHg.

#### CAUTIONS

Federal (U.S.A.) law restricts this device to sale by or on the order of a physician.

Do not immerse electrical connectors. Do not autoclave or kink cables.

#### INSTRUCTIONS FOR USE

1. Connect transducer end of MPS interface cable to the disposable transducer cable connector.
2. Plug MPS interface cable connector into the retrograde (blue) port on back of the MPS console.
3. Follow disposable pressure transducer setup procedure to ensure proper priming and zeroing.

For symbols glossary visit [www.questmedical.com/resources.aspx](http://www.questmedical.com/resources.aspx)



One Allentown Parkway, Allen, Texas 75002 USA  
+1.972.390.9800 | +1.800.627.0226 | F +1.972.390.8465

© 2019 Quest Medical, Inc. MPS is a registered trademark of Quest Medical, Inc. The MPS Console and MPS sterile disposables are covered under one or more of the following U.S. patents: 5,573,502 . 5,588,816 . 5,638,737 . 5,645,531 . 5,899,873. Also covered by pending U.S. and International Patents and patent applications.



EMERGO EUROPE  
Prinsessegracht 20  
2514 AP The Hague  
The Netherlands

**Made in USA**  
903289 REV E  
JUL 19