


REF 511800A, 5140004

1 UNIT



DESCRIPTION:

Transducer interface cable with red collared male connector. Cable connects a disposable pressure transducer with the MPS®3 console.

INDICATION FOR USE

The transducer interface cable is indicated to be used as the electronic interface between the MPS 3 console and a disposable pressure transducer for purpose of monitoring a defined pressure.

CONTRAINDICATION

This product is designed and warranted for use with the Quest MPS 3 cardioplegia delivery system only. Use with other systems is contraindicated. This product is not designed, sold, or intended for use except as indicated.

WARNINGS

Read instructions carefully. The attending clinician is solely responsible for the setup and use of this device.

Use cable to interface disposable pressure transducer with the MPS 3 console.

The transducer will not accurately measure pressures above 300 mmHg.

CAUTIONS

Federal (U.S.A.) law restricts this device to sale by or on the order of a physician.

Do not immerse electrical connectors. Do not autoclave or kink cables.

INSTRUCTIONS FOR USE

1. Connect transducer end of MPS 3 interface cable to the disposable transducer cable connector.
2. Plug MPS 3 interface cable connector into the antegrade (red) port on back of the MPS 3 console.
3. If using a 5140004 holder, ensure it is securely mounted to a pole (0.5" - 1.5" diameter). Slide pressure transducers until they snap in place.
4. Follow disposable pressure transducer setup procedure to ensure proper priming and zeroing.
5. Adjust the colored tabs to the appropriate configuration to use.
6. To reposition the holder on the pole: unclamp, adjust the position, reclamp, and ensure transducers are properly leveled.